

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5185

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2024 09:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 73.2 F 0.0  
4:00 86.3 F 0.0  
6:00 115.8 F 0.0  
8:00 149.0 F 0.1  
10:00 177.2 F 0.2  
12:00 198.4 F 0.5  
14:00 215.4 F 3.8  
16:00 222.6 F 6.6  
18:00 231.2 F 9.9  
20:00 239.9 F 11.1  
22:00 239.2 F 10.2  
24:00 237.1 F 8.4  
26:00 247.3 F 12.8  
28:00 244.9 F 12.2  
30:00 256.9 F 18.6  
32:00 267.5 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.8  
1:00 272.8 F 28.6  
2:00 271.9 F 27.8  
3:00 272.4 F 28.5  
4:00 272.5 F 28.2  
  
Min 271.6 F 27.7  
Max 272.9 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 194.4 F 0.0  
10:00 188.6 F 0.0  
15:00 187.3 F 0.0  
20:00 194.8 F 0.0  
25:00 227.0 F 0.0  
30:00 232.6 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 33:28  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:11:29  
CYCLE COMPLETE

12/13/2024 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_