
Midmark M11 - v1.0.5
Total Cycles: 111

Name: RITTER2
Sterilizer ID: V2313228
06/17/2021 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 163.0 F 2.0
4:00 167.0 F 5.0
6:00 187.2 F 8.7
8:00 206.0 F 12.9
10:00 216.6 F 5.9
12:00 228.2 F 10.2
14:00 232.3 F 8.9
16:00 241.2 F 12.3
18:00 241.2 F 11.1
20:00 252.4 F 16.9
22:00 262.5 F 22.8
24:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.9
2:00 272.1 F 29.3
3:00 271.7 F 29.4
4:00 272.6 F 29.8

Min 271.4 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 228.9 F 0.0
10:00 233.3 F 0.0
15:00 222.7 F 0.0
20:00 218.3 F 0.0
25:00 212.2 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:04:19
CYCLE COMPLETE

06/17/2021 10:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____