

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3495

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2023 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
2:00 192.6 F 1.0  
4:00 216.2 F 6.4  
6:00 232.3 F 11.9  
8:00 231.8 F 8.1  
10:00 240.9 F 12.1  
12:00 243.2 F 12.1  
14:00 241.9 F 11.1  
16:00 255.4 F 18.1  
18:00 267.2 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.4 F 28.7  
2:00 271.6 F 28.3  
3:00 272.4 F 28.7  
4:00 271.5 F 28.1  
  
Min 271.5 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 224.8 F 0.0  
10:00 226.5 F 0.0  
15:00 222.2 F 0.0  
20:00 220.3 F 0.0  
25:00 221.7 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 19:23  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

07/06/2023 12:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_