
Midmark M11 - v1.0.5
Total Cycles: 2999

Name: RITTER2
Sterilizer ID: V2313228
10/12/2023 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 85.6 F 0.0
4:00 86.2 F 0.3
6:00 110.9 F 2.1
8:00 142.9 F 5.1
10:00 172.3 F 8.7
12:00 214.0 F 12.7
14:00 216.4 F 4.4
16:00 225.9 F 7.9
18:00 235.9 F 11.9
20:00 231.5 F 8.0
22:00 242.0 F 13.3
24:00 241.5 F 11.4
26:00 253.5 F 17.6
28:00 264.2 F 24.1
30:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.8 F 29.5
2:00 271.7 F 29.6
3:00 272.2 F 29.4
4:00 273.2 F 29.8

Min 271.0 F 29.0
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 226.5 F 0.0
10:00 217.1 F 0.0
15:00 225.4 F 0.0
20:00 214.0 F 0.0
25:00 223.2 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:03
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:59
CYCLE COMPLETE

10/12/2023 09:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____