
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001515
14:52 18/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 3:46
136.4 C 225kPa 3:46
136.4 C 226kPa 4:16
136.1 C 223kPa 4:46
136.4 C 226kPa 5:16
136.2 C 223kPa 5:46
136.2 C 224kPa 6:16
136.4 C 226kPa 6:46
136.2 C 224kPa 7:16
136.2 C 224kPa 7:16
VENTING 7:16
AIR DRYING 8:07
CYCLE COMPLETE 12:04

Digital Signature #
C7F740A198CEB6FC
