
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007856
9:15 11/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:59
STERILIZING 4:10
136.0 C 220kPa 4:10
136.4 C 221kPa 4:40
136.4 C 223kPa 5:10
136.3 C 223kPa 5:40
136.4 C 222kPa 6:10
136.4 C 223kPa 6:40
136.4 C 224kPa 7:10
136.4 C 223kPa 7:40
136.4 C 223kPa 7:40
VENTING 7:40
AIR DRYING 8:37
CYCLE COMPLETE 38:37

Digital Signature #
A6341F92778A5A04
