

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2725

Name: RITTER2  
Sterilizer ID: V2313228  
08/02/2023 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.2 F 0.0  
2:00 164.4 F 0.0  
4:00 182.0 F 3.0  
6:00 201.9 F 7.5  
8:00 218.3 F 12.4  
10:00 227.0 F 7.4  
12:00 237.1 F 12.4  
14:00 235.8 F 9.7  
16:00 237.4 F 9.3  
18:00 247.9 F 14.4  
20:00 259.2 F 20.8  
22:00 269.7 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.4 F 29.4  
2:00 271.3 F 28.7  
3:00 272.5 F 29.6  
4:00 272.0 F 29.0

Min 271.2 F 28.7  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
5:00 201.4 F 0.0  
10:00 200.7 F 0.0  
15:00 197.2 F 0.0  
20:00 203.4 F 0.0  
25:00 206.6 F 0.0  
30:00 206.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:02  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:07  
CYCLE COMPLETE

08/02/2023 12:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_