
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002511
9:31 25/10/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:17
136.0 C 220kPa 4:17
136.4 C 224kPa 4:47
136.3 C 223kPa 5:17
136.4 C 223kPa 5:47
136.4 C 220kPa 6:17
136.4 C 223kPa 6:47
136.4 C 221kPa 7:17
136.4 C 223kPa 7:47
136.4 C 223kPa 7:47
VENTING 7:48
AIR DRYING 8:44
CYCLE COMPLETE 8:46

Digital Signature #
81B19272878D703C
