
Midmark M11 - v1.0.5
Total Cycles: 4318

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.2 F 0.0
2:00 155.1 F 2.1
4:00 160.9 F 5.6
6:00 184.1 F 9.9
8:00 211.9 F 6.6
10:00 223.9 F 9.1
12:00 235.3 F 10.4
14:00 239.1 F 12.1
16:00 239.1 F 10.8
18:00 252.2 F 17.4
20:00 263.9 F 24.2
22:00 271.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.1
1:00 273.0 F 30.4
2:00 272.0 F 30.6
3:00 272.4 F 30.2
4:00 272.3 F 30.7

Min 271.5 F 29.9
Max 273.2 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 203.5 F 0.0
10:00 222.3 F 0.0
15:00 226.6 F 0.0
20:00 229.2 F 0.0
25:00 224.7 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:39
CYCLE COMPLETE

01/30/2025 01:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____