
Midmark M11 - v1.0.0
Total Cycles: 4667

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2024 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.6 F 0.0
2:00 115.8 F 0.7
4:00 119.3 F 2.6
6:00 146.3 F 5.2
8:00 166.4 F 8.3
10:00 201.9 F 11.8
12:00 206.0 F 3.8
14:00 215.6 F 7.2
16:00 225.8 F 10.9
18:00 230.8 F 7.1
20:00 237.9 F 11.1
22:00 234.9 F 8.3
24:00 247.8 F 14.2
26:00 258.1 F 19.8
28:00 267.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 272.1 F 28.7
3:00 271.1 F 28.3
4:00 271.9 F 29.0

Min 271.0 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 204.9 F 0.0
10:00 226.2 F 0.0
15:00 235.2 F 0.0
20:00 225.0 F 0.0
25:00 228.3 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:33
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

06/21/2024 10:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____