
Midmark M11 - v1.0.0
Total Cycles: 2437

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 79.3 F 0.0
4:00 105.3 F 0.0
6:00 114.6 F 0.5
8:00 146.8 F 3.2
10:00 173.0 F 6.3
12:00 190.3 F 9.8
14:00 206.9 F 13.3
16:00 213.4 F 5.5
18:00 224.5 F 9.3
20:00 235.3 F 13.3
22:00 236.0 F 9.6
24:00 238.0 F 9.5
26:00 245.3 F 13.1
28:00 256.5 F 19.0
30:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.2 F 29.3
2:00 272.3 F 29.1
3:00 271.5 F 28.5
4:00 271.5 F 28.9

Min 271.1 F 28.4
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 197.0 F 0.0
10:00 195.2 F 0.0
15:00 196.4 F 0.0
20:00 196.0 F 0.0
25:00 223.1 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

10/17/2022 08:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____