
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001574
9:49 11/11/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:54
PRESSURIZING 6:26
STERILIZING 7:54
136.0 C 220kPa 7:54
136.4 C 219kPa 8:24
136.3 C 222kPa 8:54
136.4 C 221kPa 9:24
136.4 C 217kPa 9:54
136.4 C 223kPa 10:25
136.2 C 221kPa 10:55
136.4 C 218kPa 11:24
VENTING 11:25
AIR DRYING 12:26
CYCLE COMPLETE 42:26

Digital Signature #
C77E117BFE30A3A0
