
Midmark M11 - v1.0.5
Total Cycles: 1287

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 08:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 179.3 F 1.3
4:00 173.3 F 4.8
6:00 196.2 F 9.3
8:00 218.4 F 6.8
10:00 226.4 F 9.6
12:00 236.7 F 9.4
14:00 243.1 F 12.5
16:00 241.1 F 11.2
18:00 253.2 F 17.5
20:00 264.0 F 23.9
22:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.3 F 29.5
2:00 271.2 F 29.0
3:00 272.2 F 29.7
4:00 271.6 F 28.9

Min 271.2 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 224.7 F 0.0
10:00 230.2 F 0.0
15:00 226.9 F 0.0
20:00 220.8 F 0.0
25:00 218.6 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

07/13/2022 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____