
Midmark M11 - v1.0.5
Total Cycles: 1738

Name: RITTER2
Sterilizer ID: V2313228
11/16/2022 06:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.3 F 0.0
2:00 193.1 F 3.7
4:00 209.0 F 9.8
6:00 222.8 F 6.9
8:00 237.3 F 12.5
10:00 239.6 F 11.4
12:00 240.6 F 11.2
14:00 255.6 F 18.9
16:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.3 F 29.4
2:00 271.9 F 29.6
3:00 272.2 F 29.3
4:00 271.7 F 29.5

Min 271.4 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 232.0 F 0.0
10:00 244.3 F 0.0
15:00 231.6 F 0.0
20:00 228.9 F 0.0
25:00 229.1 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 17:07
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:56:39
CYCLE COMPLETE

11/16/2022 07:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____