

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2878

Name: RITTER 1  
Sterilizer ID: V2178198  
02/09/2023 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.6 F 0.0  
2:00 80.2 F 0.0  
4:00 86.2 F 0.9  
6:00 118.8 F 3.2  
8:00 144.3 F 5.8  
10:00 169.1 F 9.1  
12:00 192.3 F 13.1  
14:00 207.8 F 4.5  
16:00 221.3 F 8.1  
18:00 231.4 F 12.1  
20:00 231.2 F 7.8  
22:00 243.3 F 13.1  
24:00 241.5 F 11.1  
26:00 253.7 F 17.3  
28:00 264.3 F 23.7  
30:00 271.5 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 28.9  
2:00 271.7 F 28.8  
3:00 272.2 F 28.8  
4:00 271.9 F 28.9

Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.3 F 0.0  
5:00 202.9 F 0.0  
10:00 213.3 F 0.0  
15:00 220.0 F 0.0  
20:00 219.8 F 0.0  
25:00 219.5 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 30:00  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:02  
CYCLE COMPLETE

02/09/2023 09:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_