
Midmark M11 - v1.0.5
Total Cycles: 889

Name: RITTER2
Sterilizer ID: V2313228
03/21/2022 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.4 F 0.0
2:00 198.7 F 0.6

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 216F
CHAMBER PRESS 1.8PSI
** ITEMS NOT STERILE **

03/21/2022 11:21 AM