
Midmark M11 - v1.0.5
Total Cycles: 567

Name: RITTER2
Sterilizer ID: V2313228
11/18/2021 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.7 F 0.0
2:00 198.6 F 4.2
4:00 202.7 F 10.2
6:00 226.6 F 6.1
8:00 236.2 F 12.5
10:00 239.3 F 10.9
12:00 237.6 F 10.1
14:00 253.0 F 17.4
16:00 265.5 F 24.8
18:00 272.9 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.7
1:00 272.7 F 29.7
2:00 272.7 F 30.2
3:00 272.4 F 30.3
4:00 272.6 F 29.7

Min 271.8 F 29.6
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 230.5 F 0.0
10:00 222.9 F 0.0
15:00 208.1 F 0.0
20:00 214.1 F 0.0
25:00 218.7 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 2:58
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

11/18/2021 03:23 PM
ACCEPTED: X
REJECTED:
Operator: U02