
Midmark M11 - v1.0.0
Total Cycles: 5084

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2024 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.8 F 0.0
2:00 119.3 F 0.1
4:00 130.4 F 0.4
6:00 166.2 F 0.8
8:00 193.0 F 2.5
10:00 210.6 F 6.1
12:00 219.9 F 9.7
14:00 228.3 F 10.1
16:00 233.4 F 8.8
18:00 244.3 F 13.5
20:00 242.3 F 11.3
22:00 239.4 F 9.7
24:00 252.6 F 16.0
26:00 263.4 F 22.3
28:00 272.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.7
1:00 272.7 F 28.4
2:00 271.6 F 27.6
3:00 272.6 F 28.5
4:00 272.1 F 28.0

Min 271.5 F 27.6
Max 272.7 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 203.0 F 0.0
10:00 216.0 F 0.0
15:00 215.4 F 0.0
20:00 213.7 F 0.0
25:00 217.3 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:16
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:50
CYCLE COMPLETE

11/13/2024 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____