

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1233

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2022 09:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.1 F 0.0  
2:00 143.6 F 1.9  
4:00 157.0 F 5.6  
6:00 179.5 F 10.1  
8:00 211.1 F 5.4  
10:00 229.7 F 8.7  
12:00 239.7 F 13.1  
14:00 238.5 F 10.5  
16:00 235.6 F 8.2  
18:00 248.7 F 15.0  
20:00 260.2 F 21.3  
22:00 270.6 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 271.5 F 29.4  
2:00 273.0 F 29.6  
3:00 271.2 F 29.1  
4:00 272.5 F 29.9

Min 271.0 F 28.6  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 229.9 F 0.0  
10:00 230.9 F 0.0  
15:00 234.0 F 0.0  
20:00 228.1 F 0.0  
25:00 230.5 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 22:43  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:26  
CYCLE COMPLETE

06/28/2022 10:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_