
Midmark M11 - v1.0.0
Total Cycles: 1663

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2022 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 188.9 F 3.5
4:00 199.3 F 8.1
6:00 218.8 F 10.1
8:00 228.6 F 8.5
10:00 238.7 F 10.9
12:00 242.0 F 11.8
14:00 240.6 F 10.5
16:00 254.3 F 17.4
18:00 265.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.6 F 28.2
3:00 272.5 F 28.7
4:00 271.5 F 28.1

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 232.7 F 0.0
10:00 240.6 F 0.0
15:00 228.9 F 0.0
20:00 227.4 F 0.0
25:00 229.7 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

02/23/2022 04:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____