
Midmark M11 - v1.0.5
Total Cycles: 1286

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 07:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 72.8 F 0.0
4:00 82.1 F 0.0
6:00 141.8 F 0.0
8:00 141.7 F 2.0
10:00 163.5 F 4.9
12:00 199.0 F 8.0
14:00 225.1 F 11.7
16:00 216.6 F 4.7
18:00 224.3 F 8.7
20:00 235.4 F 12.8
22:00 233.4 F 9.4
24:00 240.5 F 10.7
26:00 243.8 F 12.7
28:00 254.3 F 18.3
30:00 264.1 F 24.2
32:00 271.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.1 F 29.5
2:00 272.6 F 30.0
3:00 271.5 F 29.3
4:00 272.1 F 29.5

Min 271.4 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 219.3 F 0.0
5:00 213.7 F 0.0
10:00 221.0 F 0.0
15:00 224.9 F 0.0
20:00 227.6 F 0.0
25:00 239.1 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:60
CYCLE COMPLETE

07/13/2022 08:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____