
Midmark M11 - v1.0.0
Total Cycles: 3736

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2023 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 170.7 F 1.8
4:00 185.7 F 5.4
6:00 204.3 F 9.4
8:00 219.6 F 10.3
10:00 228.0 F 8.6
12:00 236.1 F 13.4
14:00 238.4 F 10.7
16:00 236.2 F 9.0
18:00 249.8 F 15.4
20:00 260.9 F 21.6
22:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.6 F 29.2
2:00 271.3 F 28.4
3:00 272.5 F 29.1
4:00 271.4 F 28.4

Min 271.2 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 227.4 F 0.0
10:00 223.6 F 0.0
15:00 219.9 F 0.0
20:00 218.7 F 0.0
25:00 217.4 F 0.0
30:00 213.5 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:01:40
CYCLE COMPLETE

09/07/2023 03:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____