
Midmark M11 - v1.0.5
Total Cycles: 937

Name: RITTER2
Sterilizer ID: V2313228
03/31/2022 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 173.9 F 0.5
4:00 192.7 F 3.9
6:00 215.1 F 9.4
8:00 227.1 F 8.6
10:00 235.7 F 12.1
12:00 237.2 F 10.6
14:00 237.3 F 9.9
16:00 252.3 F 17.2
18:00 264.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.1
2:00 272.1 F 29.5
3:00 271.2 F 29.4
4:00 272.8 F 30.2

Min 271.2 F 29.1
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 219.0 F 0.0
10:00 232.4 F 0.0
15:00 230.8 F 0.0
20:00 226.7 F 0.0
25:00 226.3 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

03/31/2022 12:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____