
Midmark M11 - v1.0.5
Total Cycles: 1957

Name: RITTER2
Sterilizer ID: V2313228
01/23/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.2 F 0.0
2:00 198.9 F 0.8
4:00 207.4 F 6.3
6:00 224.9 F 11.9
8:00 231.9 F 9.2
10:00 236.7 F 9.5
12:00 241.1 F 11.1
14:00 248.1 F 14.7
16:00 261.2 F 22.1
18:00 272.9 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.4
1:00 272.6 F 30.1
2:00 272.7 F 29.8
3:00 271.7 F 29.3
4:00 273.1 F 30.2

Min 271.7 F 29.2
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 233.9 F 0.0
10:00 229.9 F 0.0
15:00 227.6 F 0.0
20:00 226.6 F 0.0
25:00 227.5 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

01/23/2023 12:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____