
Midmark M11 - v1.0.0
Total Cycles: 5428

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2025 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 71.6 F 0.4
4:00 87.4 F 2.0
6:00 124.8 F 4.5
8:00 154.2 F 7.3
10:00 177.5 F 10.8
12:00 193.7 F 3.1
14:00 210.1 F 5.5
16:00 222.5 F 9.0
18:00 232.9 F 12.9
20:00 231.3 F 8.1
22:00 243.1 F 13.2
24:00 240.5 F 10.9
26:00 253.2 F 17.2
28:00 264.3 F 23.7
30:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.0
2:00 272.2 F 29.1
3:00 272.1 F 28.6
4:00 272.7 F 29.4

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 204.4 F 0.0
10:00 210.4 F 0.0
15:00 206.0 F 0.0
20:00 206.3 F 0.0
25:00 214.3 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:01
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:08:09
CYCLE COMPLETE

03/21/2025 09:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____