

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4079

Name: RITTER2  
Sterilizer ID: V2313228  
11/04/2024 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.8 F 0.0  
2:00 161.1 F 1.8  
4:00 181.3 F 5.2  
6:00 199.0 F 8.9  
8:00 212.6 F 13.1  
10:00 219.1 F 6.8  
12:00 230.5 F 11.1  
14:00 230.2 F 7.6  
16:00 243.2 F 13.3  
18:00 243.3 F 12.2  
20:00 254.9 F 18.3  
22:00 265.4 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.8 F 29.6  
2:00 271.6 F 28.9  
3:00 272.6 F 29.8  
4:00 272.2 F 29.3

Min 271.4 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 193.9 F 0.0  
10:00 194.3 F 0.0  
15:00 189.6 F 0.0  
20:00 183.6 F 0.0  
25:00 180.1 F 0.0  
30:00 177.3 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 23:54  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:03:11  
CYCLE COMPLETE

11/04/2024 12:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_