
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000009
9:09 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:14
STERILIZING 4:04
136.0 C 220kPa 4:04
136.1 C 222kPa 4:34
136.2 C 224kPa 5:04
136.2 C 220kPa 5:34
136.3 C 224kPa 6:04
136.3 C 222kPa 6:34
136.3 C 220kPa 7:04
136.1 C 222kPa 7:34
136.1 C 222kPa 7:34
VENTING 7:34
AIR DRYING 8:22
CYCLE COMPLETE 38:22

Digital Signature #
AB21F48A3E9F108B
