
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001977
11:06 08/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:09
STERILIZING 4:05
136.0 C 220kPa 4:05
135.9 C 221kPa 4:35
136.4 C 223kPa 5:05
136.1 C 223kPa 5:35
136.4 C 225kPa 6:05
136.2 C 223kPa 6:35
136.3 C 223kPa 7:05
136.3 C 223kPa 7:35
136.3 C 223kPa 7:35
VENTING 7:36
CYCLE ABORTED 7:58
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:58

Digital Signature #
25B6570FAAF14472
