
Midmark M11 - v1.0.0
Total Cycles: 2998

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2023 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.6 F 0.0
2:00 186.0 F 3.5
4:00 206.0 F 8.3
6:00 228.7 F 10.0
8:00 233.8 F 9.9
10:00 236.2 F 8.1
12:00 243.9 F 12.4
14:00 245.1 F 12.9
16:00 258.4 F 20.0
18:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.4 F 29.2
2:00 271.9 F 28.5
3:00 271.9 F 29.0
4:00 272.4 F 28.9

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 222.5 F 0.0
10:00 223.2 F 0.0
15:00 222.0 F 0.0
20:00 231.8 F 0.0
25:00 228.6 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

03/13/2023 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____