
Midmark M11 - v1.0.0
Total Cycles: 3577

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2023 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 170.2 F 0.4
4:00 185.3 F 2.8
6:00 203.0 F 6.3
8:00 214.5 F 10.3
10:00 221.6 F 7.8
12:00 229.5 F 8.3
14:00 239.5 F 12.8
16:00 238.7 F 10.2
18:00 237.8 F 9.0
20:00 248.8 F 14.5
22:00 259.7 F 20.5
24:00 269.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.5 F 28.1
3:00 272.4 F 29.0
4:00 272.1 F 28.5

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 210.4 F 0.0
10:00 209.9 F 0.0
15:00 207.6 F 0.0
20:00 209.9 F 0.0
25:00 210.9 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

07/25/2023 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____