
Midmark M11 - v1.0.0
Total Cycles: 2123

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.3 F 0.0
2:00 168.3 F 1.2
4:00 176.3 F 5.1
6:00 214.4 F 10.2
8:00 224.4 F 5.2
10:00 231.8 F 9.4
12:00 238.5 F 9.7
14:00 239.4 F 11.3
16:00 237.3 F 9.2
18:00 250.5 F 15.7
20:00 261.6 F 22.1
22:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.4 F 29.0
2:00 271.0 F 28.3
3:00 272.4 F 29.2
4:00 271.6 F 28.4

Min 271.0 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 230.9 F 0.0
10:00 228.9 F 0.0
15:00 234.5 F 0.0
20:00 230.9 F 0.0
25:00 225.5 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

07/13/2022 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____