
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000591
14:02 19/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:30
STERILIZING 4:46
136.0 C 222kPa 4:46
136.4 C 232kPa 5:16
136.1 C 230kPa 5:46
136.4 C 231kPa 6:16
136.3 C 232kPa 6:46
136.3 C 229kPa 7:16
136.3 C 231kPa 7:46
136.0 C 230kPa 8:16
136.0 C 230kPa 8:16
VENTING 8:16
AIR DRYING 9:06
CYCLE COMPLETE 15:08

Digital Signature #
AC2EB5E842808891
