
Midmark M11 - v1.0.0
Total Cycles: 3323

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2023 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 182.5 F 4.2
4:00 199.4 F 9.1
6:00 215.5 F 7.9
8:00 224.7 F 9.1
10:00 234.4 F 9.8
12:00 241.1 F 12.1
14:00 240.4 F 10.8
16:00 254.5 F 17.7
18:00 266.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.5 F 28.6
3:00 272.4 F 29.0
4:00 271.4 F 28.3

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 221.1 F 0.0
10:00 218.7 F 0.0
15:00 218.0 F 0.0
20:00 223.1 F 0.0
25:00 225.7 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

06/01/2023 02:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____