
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006381
14:27 28/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:06
STERILIZING 3:08
136.0 C 220kPa 3:08
136.0 C 222kPa 3:38
136.1 C 222kPa 4:08
136.1 C 223kPa 4:38
136.4 C 225kPa 5:08
136.4 C 224kPa 5:38
136.4 C 221kPa 6:08
136.2 C 223kPa 6:38
136.2 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:28
CYCLE COMPLETE 7:35

Digital Signature #
31641F5EA2B3BC9E
