
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001390
11:13 19/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:46
STERILIZING 5:03
136.0 C 220kPa 5:03
136.3 C 224kPa 5:33
136.3 C 223kPa 6:03
136.1 C 222kPa 6:33
136.3 C 224kPa 7:03
136.3 C 223kPa 7:33
136.4 C 222kPa 8:03
136.3 C 224kPa 8:33
136.3 C 224kPa 8:33
VENTING 8:33
AIR DRYING 9:31
CYCLE COMPLETE 24:31

Digital Signature #
91F21B84D16986FB
