

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3485

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2023 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.9 F 0.0  
2:00 196.8 F 2.2  
4:00 206.1 F 6.2  
6:00 215.2 F 11.9  
8:00 223.7 F 6.9  
10:00 236.7 F 12.5  
12:00 237.1 F 9.7  
14:00 237.4 F 8.7  
16:00 249.3 F 14.9  
18:00 261.5 F 21.7  
20:00 272.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.4 F 29.1  
2:00 271.8 F 28.4  
3:00 272.1 F 29.0  
4:00 272.2 F 28.7  
  
Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 215.7 F 0.0  
10:00 236.4 F 0.0  
15:00 235.2 F 0.0  
20:00 228.7 F 0.0  
25:00 227.9 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 20:19  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:02  
CYCLE COMPLETE

07/05/2023 09:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_