

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 003656  
15:26 11/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:04  
STERILIZING 3:03  
136.4 C 221kPa 3:03  
136.4 C 221kPa 3:33  
136.4 C 221kPa 4:03  
136.4 C 225kPa 4:33  
136.0 C 222kPa 5:03  
136.2 C 224kPa 5:33  
136.2 C 224kPa 6:03  
136.3 C 224kPa 6:33  
136.3 C 224kPa 6:33  
VENTING 6:33  
AIR DRYING 7:25  
CYCLE COMPLETE 7:36

-----  
Digital Signature #  
2A627FD22DFBA146  
-----