
Midmark M11 - v1.0.0
Total Cycles: 1943

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2022 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.2 F 0.0
2:00 136.3 F 1.8
4:00 149.8 F 4.6
6:00 173.7 F 8.1
8:00 191.2 F 12.4
10:00 213.8 F 5.1
12:00 226.4 F 9.7
14:00 232.4 F 8.7
16:00 238.1 F 11.0
18:00 235.0 F 8.4
20:00 248.7 F 14.7
22:00 259.9 F 21.0
24:00 270.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.1 F 28.7
2:00 272.5 F 29.3
3:00 271.4 F 28.5
4:00 272.1 F 28.9

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 228.3 F 0.0
10:00 225.9 F 0.0
15:00 224.1 F 0.0
20:00 227.5 F 0.0
25:00 241.7 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

05/18/2022 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____