
Midmark M11 - v1.0.0
Total Cycles: 3291

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2023 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.1 F 0.0
2:00 158.2 F 0.0
4:00 182.2 F 2.4
6:00 198.1 F 6.5
8:00 214.7 F 11.0
10:00 220.8 F 5.1
12:00 234.2 F 10.2
14:00 234.8 F 8.4
16:00 245.1 F 13.1
18:00 243.7 F 11.8
20:00 256.2 F 18.3
22:00 267.0 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.4 F 28.8
2:00 272.1 F 28.2
3:00 271.7 F 28.4
4:00 272.4 F 28.6

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 214.8 F 0.0
10:00 214.8 F 0.0
15:00 207.9 F 0.0
20:00 201.9 F 0.0
25:00 199.4 F 0.0
30:00 202.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:24
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:43
CYCLE COMPLETE

05/26/2023 02:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____