
Midmark M11 - v1.0.5
Total Cycles: 4213

Name: RITTER2
Sterilizer ID: V2313228
12/13/2024 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.9 F 0.0
2:00 121.5 F 0.7
4:00 139.7 F 3.1
6:00 165.5 F 6.3
8:00 184.8 F 10.2
10:00 217.9 F 7.1
12:00 223.8 F 7.3
14:00 230.0 F 11.8
16:00 233.3 F 7.9
18:00 242.7 F 13.1
20:00 241.9 F 11.4
22:00 253.5 F 17.2
24:00 264.0 F 23.5
26:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 273.2 F 29.5
2:00 271.5 F 29.2
3:00 272.5 F 29.5
4:00 271.5 F 29.0

Min 271.5 F 28.9
Max 273.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 238.1 F 0.0
10:00 226.5 F 0.0
15:00 232.2 F 0.0
20:00 215.6 F 0.0
25:00 221.8 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:58
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:36
CYCLE COMPLETE

12/13/2024 04:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____