
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001536
13:44 10/27/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 5:39
STERILIZING 6:56
136.0 C 220kPa 6:56
136.3 C 221kPa 7:26
136.3 C 222kPa 7:56
136.3 C 221kPa 8:26
136.5 C 222kPa 8:56
136.5 C 220kPa 9:26
136.3 C 222kPa 9:56
136.0 C 220kPa 10:26
VENTING 10:26
AIR DRYING 11:30
CYCLE COMPLETE 41:30

Digital Signature #
49F2245B5D78CC45
