

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3568

Name: RITTER2  
Sterilizer ID: V2313228  
04/29/2024 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.1 F 0.0  
2:00 169.0 F 1.9  
4:00 168.7 F 5.1  
6:00 187.8 F 9.2  
8:00 205.7 F 12.1  
10:00 226.3 F 7.4  
12:00 229.3 F 12.2  
14:00 234.6 F 9.1  
16:00 239.8 F 10.4  
18:00 246.2 F 13.7  
20:00 257.9 F 20.0  
22:00 268.3 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 271.8 F 29.6  
2:00 273.0 F 29.9  
3:00 271.6 F 29.5  
4:00 272.5 F 30.1

Min 270.9 F 28.7  
Max 273.4 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
5:00 231.1 F 0.0  
10:00 214.8 F 0.0  
15:00 185.1 F 0.0  
20:00 182.9 F 0.0  
25:00 182.0 F 0.0  
30:00 180.4 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 23:12  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:01:55  
CYCLE COMPLETE

04/29/2024 11:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_