

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1471

Name: RITTER 1  
Sterilizer ID: V2178198  
12/16/2021 04:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
2:00 207.0 F 5.1  
4:00 222.5 F 12.0  
6:00 231.4 F 8.8  
8:00 237.3 F 9.3  
10:00 243.6 F 11.8  
12:00 246.8 F 13.7  
14:00 260.6 F 21.2  
16:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.2 F 28.8  
2:00 271.9 F 28.8  
3:00 272.1 F 28.7  
4:00 271.8 F 28.7  
  
Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 226.5 F 0.0  
10:00 227.5 F 0.0  
15:00 228.3 F 0.0  
20:00 230.1 F 0.0  
25:00 228.6 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 3:34  
HEATING: 16:24  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:56:59  
CYCLE COMPLETE

12/16/2021 05:06 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02