
Midmark M11 - v1.0.5
Total Cycles: 775

Name: RITTER2
Sterilizer ID: V2313228
02/09/2022 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 76.5 F 0.0
4:00 87.6 F 0.0
6:00 110.3 F 0.1
8:00 151.4 F 2.2
10:00 175.4 F 5.1
12:00 194.7 F 8.5
14:00 206.0 F 12.7
16:00 218.4 F 6.2
18:00 229.4 F 10.3
20:00 233.5 F 8.9
22:00 241.6 F 12.5
24:00 240.9 F 11.3
26:00 253.3 F 17.6
28:00 264.4 F 24.1
30:00 271.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.8
2:00 271.8 F 28.9
3:00 272.1 F 29.7
4:00 272.5 F 29.5

Min 271.4 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 197.4 F 0.0
10:00 204.1 F 0.0
15:00 195.5 F 0.0
20:00 195.8 F 0.0
25:00 195.2 F 0.0
30:00 197.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 30:04
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

02/09/2022 03:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____