
Midmark M11 - v1.0.0
Total Cycles: 3699

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2023 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.8 F 0.0
2:00 173.5 F 3.7
4:00 192.6 F 8.0
6:00 212.6 F 13.0
8:00 223.2 F 7.6
10:00 235.6 F 12.5
12:00 235.4 F 9.6
14:00 237.4 F 9.3
16:00 247.3 F 14.0
18:00 258.9 F 20.2
20:00 269.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.0
2:00 271.4 F 28.4
3:00 272.4 F 29.0
4:00 271.4 F 28.2

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 207.2 F 0.0
10:00 215.5 F 0.0
15:00 217.4 F 0.0
20:00 213.7 F 0.0
25:00 210.3 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

08/29/2023 02:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____