

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4820

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2024 08:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.0 F 0.0  
2:00 77.7 F 0.0  
4:00 91.1 F 0.0  
6:00 125.3 F 0.4  
8:00 148.8 F 1.2  
10:00 177.8 F 2.8  
12:00 196.8 F 5.5  
14:00 210.0 F 9.4  
16:00 220.8 F 13.0  
18:00 222.9 F 6.3  
20:00 233.8 F 10.6  
22:00 233.2 F 7.6  
24:00 242.9 F 12.1  
26:00 239.5 F 10.0  
28:00 252.7 F 16.5  
30:00 263.8 F 23.0  
32:00 272.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 272.1 F 28.5  
2:00 272.4 F 28.5  
3:00 271.5 F 28.1  
4:00 272.7 F 28.9  
  
Min 271.4 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.3 F 0.0  
5:00 209.0 F 0.0  
10:00 201.2 F 0.0  
15:00 210.7 F 0.0  
20:00 215.3 F 0.0  
25:00 211.3 F 0.0  
30:00 213.2 F 0.0  
  
Drying Complete

FILLING: 0:40  
HEATING: 32:02  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:04  
CYCLE COMPLETE

08/15/2024 09:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_