

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2157

Name: RITTER2  
Sterilizer ID: V2313228  
03/16/2023 01:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.0 F 0.0  
2:00 134.7 F 1.9  
4:00 153.2 F 4.7  
6:00 179.5 F 9.2  
8:00 208.2 F 7.9  
10:00 221.4 F 8.1  
12:00 234.3 F 13.1  
14:00 235.9 F 10.1  
16:00 235.7 F 8.6  
18:00 248.6 F 15.1  
20:00 260.5 F 21.9  
22:00 271.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.3  
1:00 272.6 F 30.0  
2:00 271.6 F 29.2  
3:00 272.9 F 30.0  
4:00 271.8 F 29.4

Min 271.3 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 211.1 F 0.0  
10:00 212.4 F 0.0  
15:00 215.8 F 0.0  
20:00 227.5 F 0.0  
25:00 230.2 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:29  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:51  
CYCLE COMPLETE

03/16/2023 02:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_