

-----  
Midmark M11 - v1.0.5  
Total Cycles: 581

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2021 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
2:00 170.6 F 3.1  
4:00 180.5 F 7.4  
6:00 203.2 F 12.7  
8:00 223.0 F 6.6  
10:00 235.6 F 11.9  
12:00 234.5 F 9.3  
14:00 238.9 F 9.8  
16:00 247.8 F 14.6  
18:00 260.0 F 21.5  
20:00 271.0 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.1  
1:00 272.6 F 29.9  
2:00 271.9 F 29.0  
3:00 272.5 F 29.9  
4:00 272.1 F 29.2  
  
Min 271.4 F 29.0  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 221.2 F 0.0  
10:00 210.2 F 0.0  
15:00 213.7 F 0.0  
20:00 207.0 F 0.0  
25:00 212.9 F 0.0  
30:00 213.0 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 20:35  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

11/29/2021 12:40 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02