
Midmark M11 - v1.0.0
Total Cycles: 704

Name: MIDMARK
Sterilizer ID: V2178198
02/25/2021 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.1 F 0.0
2:00 114.7 F 0.1
4:00 137.2 F 0.1
6:00 171.0 F 3.1
8:00 194.0 F 7.5
10:00 214.5 F 12.1
12:00 220.8 F 6.8
14:00 233.5 F 12.1
16:00 235.8 F 9.1
18:00 238.7 F 9.4
20:00 247.1 F 13.6
22:00 259.4 F 20.4
24:00 270.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 271.7 F 28.3
2:00 272.4 F 28.9
3:00 272.1 F 29.0
4:00 272.4 F 29.4

Min 271.0 F 28.1
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 236.4 F 0.0
10:00 232.5 F 0.0
15:00 232.8 F 0.0
20:00 224.2 F 0.0
25:00 216.6 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:41
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

02/25/2021 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____