

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3053

Name: RITTER 1  
Sterilizer ID: V2178198  
03/29/2023 11:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.3 F 0.0  
2:00 173.1 F 3.6  
4:00 184.8 F 7.9  
6:00 208.3 F 12.8  
8:00 220.6 F 7.1  
10:00 234.0 F 12.4  
12:00 235.4 F 9.5  
14:00 237.3 F 8.8  
16:00 248.9 F 14.8  
18:00 261.2 F 21.8  
20:00 272.5 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 272.1 F 29.3  
2:00 272.5 F 29.1  
3:00 271.4 F 28.5  
4:00 272.5 F 29.5

Min 271.4 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 222.8 F 0.0  
10:00 223.7 F 0.0  
15:00 224.3 F 0.0  
20:00 222.8 F 0.0  
25:00 223.7 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

03/29/2023 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_